

# KU-11B

● Urine Analyzer



# KU-11B

● Urine Analyzer



## Specification

**Test item:** GLU, BIL, SG, PH, KET, BLD, PRO, URO, NIT, LEU, VC

**Wavelength:** 470nm, 540nm, 620nm

**Repeatability CV**≤1%

**Stability:** within 8 hours working, CV≤1%

**Test throughput:** 120 tests /hour or 60 tests /hour

**Language:** English

**Display:** Large LCD display, Resolution 240X128

**Printer:** Built-in thermal recorder, paper width 57mm

**Interface:** Standard RS-232

**Storage:** 1000 results

**Light source:** Super-high brightness LED white light source, with 30,000 hours lifespan

**Operating environment:** The instrument should be installed on a stable and solid platform, Avoid direct sunlight and away from strong electric and magnetic field interference.

**Temperature:** 15°C-30°C

**Humidity:** ≤ 80%

**Atmospheric pressure:** 76-106kPa

**Power supply:** 100-120VAC/200-240VAC, 50Hz/60Hz 40VA

**Dimension:** 253X237X171mm



**Pioway Medical Lab Equipment Co., Ltd.**