

KU-11C

● Urine Analyzer



KU-11C

● Urine Analyzer



Specification

Test item: GLU, BIL, SG, PH, KET, BLD, PRO, URO, NIT, LEU, VC
Wavelength: 470nm, 540nm, 620nm
Repeatability CV≤1%
Stability: within 8 hours working, CV≤1%
Test throughput: 500 tests /hour
Language: English
Display: Large LCD touch screen, Resolution 240X128
Printer: Built-in thermal recorder, paper width 57mm
Interface: Standard RS-232
Storage: 2000 results
Light source: Super-high brightness LED white light source, with 30,000 hours lifespan

Operating environment: The instrument should be installed on a stable and solid platform, Avoid direct sunlight and away from strong electric and magnetic field interference.

Temperature: 15℃-30℃

Humidity: ≤ 80%

Atmospheric pressure: 76-106kPa

Power supply: 100-120VAC/200-240VAC, 50Hz/60Hz 40VA

Dimension: 390X330X210mm



Pioway Medical Lab Equipment Co., Ltd.